



1.0 Requirements

- 1.1 The supplier shall provide calibration services that are fully compliant with one of the following requirements (latest revision):
 - 1.1.1 ANSI/NCSL.Z540.1
 - 1.1.2 ISO/IEC 17025
 - 1.1.3 ANSI/NCSL Z540.3
- 1.2 The supplier shall provide certificates of calibration that contain the information required by the appropriate standard including calibration results data upon request to Tampa Brass & Aluminum Corp.
- 1.3 The supplier shall notify Tampa Brass & Aluminum Corp. when an out of tolerance indication is discovered. The supplier shall provide information on the extent and scope of the out of tolerance indication as part of this notification.

DOCUMENT IS UNCONTROLLED WHEN PRINTED